

**SPECIAL INSTRUCTIONS
FOR TLD IRRADIATION WITH A TOMOTHERAPY UNIT**

1. Open the collimator to a 40 cm X 5 cm field.
2. Set the SSD (85 cm) to the top of the platform.
3. Set the time to deliver 300 cGy to the reference point without machine rotation (stationary beam).